

Our Ref: M03035697 an
Purchase Order No. 23216

3rd October 2003

Dominant (Australia) Pty Ltd
P O Box 306
WELLAND SA 5007

ATTENTION: MR ODISSEAS KEHAGIAS

Dear Sir,

RE: PRESERVATIVE EFFICACY OF DOMINANT 4 CONTROL

One sample marked 'Dominant 4 Control 0.3% Triclosan 14/8/03' was evaluated in accordance with the BP (2002) test for preservative efficacy (Appendix XVIC).

The following results were obtained:

Organism	Initial Count per g	Final Count per g			
		48 hrs	7 days	14 days	28 days
<i>Ps.aeruginosa</i> ATCC 9027	1.9 x 10 ⁶	<10	<10	<10	<10
<i>S.aureus</i> ATCC 6538	5.0 x 10 ⁶	<100	<100	<100	<100
<i>C.albicans</i> ATCC 10231	7.4 x 10 ⁵	-	-	<10	<10
<i>A.niger</i> ATCC 16404	5.0 x 10 ⁵	-	-	<10	<10

Notes: 1. '<' indicates Less Than.

2. Preservative inactivation and controls conformed to the requirements of the test.

The sample tested was found **to meet** the pass criteria of the test for topical preparations.

Yours faithfully,



DERIO COMAR BSc(HONS),FRACI,AAIFST,MASM
SENIOR CONSULTANT MICROBIOLOGIST
DIRECTOR

The data pertains solely to the analytical and sampling procedure(s) used and the condition and homogeneity of the sample(s) as received. The data therefore may not be representative of the lot or batch or other samples. Consequently the data may not necessarily justify the acceptance or rejection of a lot or batch, a product recall or support legal proceedings. It is the responsibility of the client to provide all information relevant to the analysis requested. This report does not imply that Silliker Microtech Pty Ltd has been engaged to consult upon the consequences of the analysis and for any action that should be taken as a result of the analysis.



NATA Accredited Laboratory Numbers 2020 and 2863
This laboratory is accredited by the National
Association of Testing Authorities, Australia.
The test(s) reported herein have been performed in
accordance with its terms of accreditation.
This document shall not be reproduced except in full.

SILLIKER MICROTECH PTY LTD
ABN 94 006 462 335
20 KING STREET
BLACKBURN, VICTORIA 3130
TEL. +61 3 9877 8222
FAX. +61 3 9877 8444
EMAIL sales@microtechlab.com
SYDNEY:
TEL. +61 2 8718 6888 FAX. +61 2 8718 6899
WWW.MICROTECHLAB.COM • WWW.SILLIKER.COM